

ISE Calibrator indirect/Urine

Version 1.3

Revision Date 09-19-2014

Print Date 12-06-2014

SECTION 1. IDENTIFICATION

Product name : ISE Calibrator indirect/Urine

Mat.-No./ Genisys-No. : 04838149001

Manufacturer or supplier's detailsCompany name of supplier : Roche Diagnostics
-Address : 9115 Hague Road
46250 Indianapolis IN

Telephone : 1-800-428-5074

Emergency telephone number:

In case of emergencies: : CHEMTREC 1-800-424-9300 (U.S. or
Canada)
1-703-527-3887 (International)**Recommended use of the chemical and restrictions on use**

Restrictions on use : For professional users only.

SECTION 2. HAZARDS IDENTIFICATION**GHS Classification**

The product is a kit consisting of individual ingredients. The classification of the ingredients can be obtained from section 3. Section Label elements contains the resulting labelling for the kit.

GHS Label element

Not a hazardous substance or mixture.

Other hazards

None known.

SECTION 3. COMPOSITION/INFORMATION ON INGREDIENTS***R1 ISE Calibrator Indirect*****GHS Classification**

Not a hazardous substance or mixture.

Hazardous components

No hazardous ingredients

SECTION 4. FIRST AID MEASURES

General advice : Do not leave the victim unattended.

If inhaled : Move to fresh air.
If unconscious place in recovery position and seek medical advice.
If symptoms persist, call a physician.

ISE Calibrator indirect/Urine

Version 1.3

Revision Date 09-19-2014

Print Date 12-06-2014

In case of skin contact	: If on skin, rinse well with water.
In case of eye contact	: Immediately flush eye(s) with plenty of water. Remove contact lenses. Protect unharmed eye. If eye irritation persists, consult a specialist.
If swallowed	: Keep respiratory tract clear. Do not give milk or alcoholic beverages. Never give anything by mouth to an unconscious person. If symptoms persist, call a physician. Rinse mouth with water.
Most important symptoms and effects, both acute and delayed	: No information available.
Notes to physician	: The first aid procedure should be established in consultation with the doctor responsible for industrial medicine.

SECTION 5. FIREFIGHTING MEASURES

Suitable extinguishing media	: Use extinguishing measures that are appropriate to local circumstances and the surrounding environment.
Specific hazards during firefighting	: No information available.
Further information	: Standard procedure for chemical fires. Use extinguishing measures that are appropriate to local circumstances and the surrounding environment.
Special protective equipment for firefighters	: Wear self-contained breathing apparatus for firefighting if necessary.

SECTION 6. ACCIDENTAL RELEASE MEASURES

Personal precautions, protective equipment and emergency procedures	: Refer to protective measures listed in sections 7 and 8.
Environmental precautions	: Local authorities should be advised if significant spillages cannot be contained.
Methods and materials for containment and cleaning up	: Wipe up with absorbent material (e.g. cloth, fleece). Keep in suitable, closed containers for disposal.

SECTION 7. HANDLING AND STORAGE

Advice on safe handling	: For personal protection see section 8. Smoking, eating and drinking should be prohibited in the application area.
Conditions for safe storage	: Electrical installations / working materials must comply with

ISE Calibrator indirect/Urine

Version 1.3

Revision Date 09-19-2014

Print Date 12-06-2014

the technological safety standards.

Materials to avoid : No materials to be especially mentioned.

SECTION 8. EXPOSURE CONTROLS/PERSONAL PROTECTION***R1 ISE Calibrator Indirect*****Components with workplace control parameters**

Contains no substances with occupational exposure limit values.

Personal protective equipment

Respiratory protection : No personal respiratory protective equipment normally required.

Hand protection

Material : Protective gloves
Remarks : The selected protective gloves have to satisfy the specifications of EU Directive 89/686/EEC and the standard EN 374 derived from it. This recommendation is only valid for the product mentioned in the safety data sheet and provided by us and for the application specified by us. Please observe the instructions regarding permeability and breakthrough time which are provided by the supplier of the gloves. Also take into consideration the specific local conditions under which the product is used, such as the danger of cuts, abrasion, and the contact time. The suitability for a specific workplace should be discussed with the producers of the protective gloves.

Eye protection : Safety glasses

Skin and body protection : Protective suit

Hygiene measures : Handle in accordance with good industrial hygiene and safety practice.

SECTION 9. PHYSICAL AND CHEMICAL PROPERTIES***R1 ISE Calibrator Indirect***

Appearance : liquid

Colour : colourless

pH : 7.4

Melting point/range : 0 °C

Boiling point/boiling range : 100 °C

Flash point : does not flash

Flammability (solid, gas) : The product is not flammable.

Upper explosion limit : No data available

ISE Calibrator indirect/Urine

Version 1.3

Revision Date 09-19-2014

Print Date 12-06-2014

Lower explosion limit	: No data available
Density	: 0.9992 g/cm3
Solubility(ies)	
Water solubility	: completely miscible
Auto-ignition temperature	: No data available
Thermal decomposition	: No data available
Oxidizing properties	: The substance or mixture is not classified as oxidizing.

SECTION 10. STABILITY AND REACTIVITY

Reactivity	: No dangerous reaction known under conditions of normal use.
Chemical stability	: Stable under normal conditions.
Possibility of hazardous reactions	: Stable under recommended storage conditions. No hazards to be specially mentioned.
Conditions to avoid	: No data available
Incompatible materials	: No data available
Hazardous decomposition products	: No data available

SECTION 11. TOXICOLOGICAL INFORMATION***R1 ISE Calibrator Indirect*****Information on likely routes of exposure****Acute toxicity**

Not classified based on available information.

Product:

Acute oral toxicity : Acute toxicity estimate : > 5,000 mg/kg
Method: Calculation method

Skin corrosion/irritation

Not classified based on available information.

Serious eye damage/eye irritation

Not classified based on available information.

Respiratory or skin sensitisation

Skin sensitisation: Not classified based on available information.

Respiratory sensitisation: Not classified based on available information.

Germ cell mutagenicity

Not classified based on available information.

Carcinogenicity

ISE Calibrator indirect/Urine

Version 1.3

Revision Date 09-19-2014

Print Date 12-06-2014

Not classified based on available information.

IARC	No component of this product present at levels greater than or equal to 0.1% is identified as probable, possible or confirmed human carcinogen by IARC.
ACGIH	No component of this product present at levels greater than or equal to 0.1% is identified as a carcinogen or potential carcinogen by ACGIH.
OSHA	No component of this product present at levels greater than or equal to 0.1% is identified as a carcinogen or potential carcinogen by OSHA.
NTP	No component of this product present at levels greater than or equal to 0.1% is identified as a known or anticipated carcinogen by NTP.

Reproductive toxicity

Not classified based on available information.

STOT - single exposure

Not classified based on available information.

STOT - repeated exposure

Not classified based on available information.

Aspiration toxicity

Not classified based on available information.

SECTION 12. ECOLOGICAL INFORMATION***R1 ISE Calibrator Indirect*****Ecotoxicity****Product:**

Ecotoxicology Assessment

Toxicity Data on Soil : Not expected to adsorb on soil.

Other organisms relevant to the environment : No data available

Persistence and degradability

No data available

Bioaccumulative potential

No data available

Mobility in soil

No data available

Other adverse effects

No data available

Product:

Regulation

40 CFR Protection of Environment; Part 82 Protection of

ISE Calibrator indirect/Urine

Version 1.3

Revision Date 09-19-2014

Print Date 12-06-2014

Stratospheric Ozone - CAA Section 602 Class I Substances

Remarks

This product neither contains, nor was manufactured with a Class I or Class II ODS as defined by the U.S. Clean Air Act Section 602 (40 CFR 82, Subpt. A, App.A + B).

SECTION 13. DISPOSAL CONSIDERATIONS**Disposal methods**

- Waste from residues : Can be disposed as waste water, when in compliance with local regulations.
- Contaminated packaging : Empty containers should be taken to an approved waste handling site for recycling or disposal.
Do not re-use empty containers.

SECTION 14. TRANSPORT INFORMATION**International Regulation****IATA-DGR**

Not regulated as a dangerous good

IMDG-Code

Not regulated as a dangerous good

Transport in bulk according to Annex II of MARPOL 73/78 and the IBC Code

Not applicable for product as supplied.

National Regulations**49 CFR**

Not regulated as a dangerous good

SECTION 15. REGULATORY INFORMATION***R1 ISE Calibrator Indirect*****EPCRA - Emergency Planning and Community Right-to-Know Act****CERCLA Reportable Quantity**

Components	CAS-No.	Component RQ (lbs)	Calculated product RQ (lbs)
Acetic acid	64-19-7	5000	*

*: Calculated RQ exceeds reasonably attainable upper limit.

SARA 304 Extremely Hazardous Substances Reportable Quantity

This material does not contain any components with a section 304 EHS RQ.

SARA 302 : No chemicals in this material are subject to the reporting requirements of SARA Title III, Section 302.

ISE Calibrator indirect/Urine

Version 1.3

Revision Date 09-19-2014

Print Date 12-06-2014

SARA 313

: This material does not contain any chemical components with known CAS numbers that exceed the threshold (De Minimis) reporting levels established by SARA Title III, Section 313.

Clean Air Act

This product does not contain any hazardous air pollutants (HAP), as defined by the U.S. Clean Air Act Section 12 (40 CFR 61).

This product does not contain any chemicals listed under the U.S. Clean Air Act Section 112(r) for Accidental Release Prevention (40 CFR 68.130, Subpart F).

This product does not contain any chemicals listed under the U.S. Clean Air Act Section 111 SOCM Intermediate or Final VOC's (40 CFR 60.489).

Clean Water Act

The following Hazardous Substances are listed under the U.S. CleanWater Act, Section 311, Table 116.4A:

acetic acid	64-19-7	0.0011 %
-------------	---------	----------

The following Hazardous Chemicals are listed under the U.S. CleanWater Act, Section 311, Table 117.3:

acetic acid	64-19-7	0.0011 %
-------------	---------	----------

This product does not contain any toxic pollutants listed under the U.S. Clean Water Act Section 307

Massachusetts Right To Know

No components are subject to the Massachusetts Right to Know Act.

Pennsylvania Right To Know


water	7732-18-5	90 - 100 %
-------	-----------	------------

New Jersey Right To Know

water	7732-18-5	90 - 100 %
-------	-----------	------------

California Prop 65

: This product does not contain any chemicals known to State of California to cause cancer, birth defects, or any other reproductive harm.

SECTION 16. OTHER INFORMATION**Further information****R1 ISE Calibrator Indirect****NFPA:****HMIS III:**

HEALTH	0
FLAMMABILITY	0
PHYSICAL HAZARD	0

0 = not significant, 1 =Slight,
2 = Moderate, 3 = High
4 = Extreme, * = Chronic

**ISE Calibrator indirect/Urine**

Version 1.3

Revision Date 09-19-2014

Print Date 12-06-2014

Revision Date : 09-19-2014

The information provided in this Safety Data Sheet is correct to the best of our knowledge, information and belief at the date of its publication. The information given is designed only as a guidance for safe handling, use, processing, storage, transportation, disposal and release and is not to be considered a warranty or quality specification. The information relates only to the specific material designated and may not be valid for such material used in combination with any other materials or in any process, unless specified in the text.